



# State of New Hampshire

## Department of Safety

Division of Fire Standards and Training & Emergency Medical Services  
 98 Smokey Bear Boulevard, Concord, New Hampshire  
 Mailing Address: 33 Hazen Drive, Concord, New Hampshire 03305-0002

Business: (603) 223-4200 Fax: (603) 271-1091



Robert L. Quinn  
 Commissioner

Justin A. Cutting  
 Director

# CLINICAL BULLETIN

Bulletin #	Title			Date Issued
77	Defibrillator Pad Incompatibility			February 9, 2022
Superseded	Released By	Approved By	Source	For More Information
	FSTEMS	Chief Romanello	NASEMSO	FST/EMS (603) 223-4200

The NASEMSO Pediatric Emergency Care (PEC) Council recently learned of several issues regarding automatic external defibrillator (AED) pad incompatibility, with multi-function defibrillator and AED pads not being used in accordance with manufacturer’s recommendations, particularly in pediatric applications.

Please find the full bulletin at: <https://nasemso.org/wp-content/uploads/Defibrillator-Pad-Incompatibility-Memo-2022-0125.pdf>