



State of New Hampshire Department of Safety

John J. Barthelmes, Commissioner



Division of Fire Standards and Training and Emergency Medical Services
Office of the EMS Bureau Chief
Nick A. Mercuri, Bureau Chief

98 Smokey Bear Blvd, Concord, New Hampshire
Mailing Address: 33 Hazen Drive, Concord, New Hampshire 03305-0002

CLINICAL BULLETIN

Bulletin #	Title			Date Issued
27	Counterfeit Combat Application Tourniquet (C-A-T)			Monday July 20, 2015
Superseded	Released By	Approved By	Source	
	N. Mercuri	Dr. Jim Suozzi		

ALERT!

Recently a New Hampshire provider experienced a catastrophic failure of what is believed to be a *counterfeit* Combat Application Tourniquet (C-A-T) device. When applying the device the windless failed by snapping in half. A second device was attempted and again the windless failed. The provider was able to eventually control the hemorrhage through other improvised equipment.

The NH Bureau of Emergency Medical Services has reached out to North American Rescue Inc., the manufacturer of the Combat Application Tourniquet (C-A-T) device, and learned they have identified at least 6 (six) counterfeit devices for sale.

Here is North American Rescue’s warning:

WARNING: *There are counterfeit medical devices on the market which resemble the Combat Application Tourniquet® (C-A-T®) and are illegally using the C-A-T® trademarks – including online vendors on eBay® and Amazon®. Some of those counterfeits have catastrophically failed during actual life-saving applications.*

If you have purchased a Combat Application Tourniquet (C-A-T) please ensure you purchased it from North American Rescue or one of their authorized dealers.

Emergency Medical Services – Fire Training and Certification – Fire Academy

Business: (603) 223-4200

Fax: (603) 271-4567

Toll Free: 1-800-371-4503

TDD Access: 1-800-735-2964

<http://www.nh.gov/safety/divisions/fstems/index.html>