



# State of New Hampshire

## Department of Safety

Division of Fire Standards and Training & Emergency Medical Services  
98 Smokey Bear Boulevard, Concord, New Hampshire  
Mailing Address: 33 Hazen Drive, Concord, New Hampshire 03305-0002

Business: (603) 223-4200 Fax: (603) 271-1091



Robert L. Quinn  
Commissioner

Deborah A. Pendergast  
Director

## CLINICAL BULLETIN

Bulletin #	Title			Date Issued
72	BodyGuard Infusion Pump Recall			September 22, 2020
Superseded	Released By	Approved By	Source	For More Information
	FSTEMS	Dr. Joey Scollan, DO	FDA	FST/EMS (603) 223-4200

We have become aware of a recent updated infusion pump recall regarding the BodyGuard Infusion Pump System from CME America. It identifies a risk of over and under infusion. The FDA has identified this as a Class I recall, the most serious type of recall. Use of these devices may cause serious injury or death.

You can find the FDA link to this recall at: <https://www.bd.com/en-us/support/recall-notifications/recall-notification---cme-america-bodyguard-infusion-pump-system>

Additionally, you can find more information from CME American's website at: <https://www.bd.com/en-us/support/recall-notifications/recall-notification---cme-america-bodyguard-infusion-pump-system>